Chromium (VI) HR and LR Portable Photometers

• Advanced LED optical system

- Innovative optical design that utilizes a reference detector and focusing lens to eliminate errors from changes in the light source and from imperfections in the glass cuvette.
- LEDs have a much higher luminous efficiency, providing more light while using less power. They also produce little heat, which could otherwise affect electronic stability.

• CAL Check™

- Validate instrument performance at any time using CAL Check cuvettes made with NIST traceable standards. The CAL Check screen guides the user step-by-step through the validation process and user calibration.
- On-screen tutorial mode with animations
 - Guides users step-by-step through the measurement process
- Waterproof and floating IP67 case
- Unit of measure is displayed along with reading
- Built-in timer
 - Built-in reaction timer that ensures consistency between tests.
- Error messages on display
 - Alerts to problems including no cap, high zero, and standard too low
- GLP data
 - Displays the last calibration date.
- Auto logging
- Battery status indicator
- Auto-shut off

Significance of Use

Hexavalent chromium salts are used in various industrial applications, such as in the manufacture of paints, dyes, explosives, and ceramics, and extensively in the metal finishing and plating industries. Due to its toxicity to humans, animals, and aquatic life, hexavalent chromium is actively monitored and neutralized in wastewater from the above industries.



| Specifications | | HI97749 Chromium (VI) LR | HI97723 Chromium (VI) HR |
|------------------------------|---|---|----------------------------------|
| Measurement | Range | 0 to 300 µg/L (as Cr (VI)) | 0 to 1000 µg/L (ppb) (as Cr(VI)) |
| | Resolution | 1 µg/L | 1 µg/L |
| | Accuracy @25°C (77°F) | ±10 µg/L ±4% of reading | ±5 µg/L ±4% of reading |
| | Method | adaptation of the ASTM Manual of Water and Environmental Technology, D1687 Diphenylcarbohydrazide Method | |
| Measurement System | Light Source | light emitting diode | |
| | Bandpass filter | 525 nm | |
| | Bandpass filter bandwidth | 8 nm | |
| | Bandpass filter wavelength accuracy | ±1.0 nm | |
| | Light Detector | silicon photocell | |
| | Cuvette type | round 24.6 mm diameter (22 mm inside) | |
| Additional Specifications | Auto logging | 50 readings | |
| | Display | 128 x 64 pixel B/W LCD with backlight | |
| | Auto-off | after 15 minutes of inactivity (30 minutes before a READ measurement) | |
| | Battery type / Life | alkaline 1.5 V AA (3) / > 800 measurements (without backlight) | |
| | Environment | 0 to 50°C (32 to 122°F); 0 to 100% RH, non-serviceable | |
| | Dimensions | 142.5 x 102.5 x 50.5 mm (5.6 x 4.0 x 2.0") | |
| | Weight | 380 g (13.4 oz.) | |
| Ordering Information | HI97723 and HI97749 is supplied with sample cuvettes (2), sample caps (2), plastic stoppers (2), 1.5V AA batteries (3), instrument quality certificate, and instruction manual. CAL Check standards and testing reagents sold separately | | |
| | HI97749C includes photometer, CAL Check standards, sample cuvettes (2), sample caps (2), plastic stoppers (2), 1.5V AA batteries (3), cuvette wiping cloth, scissors, CAL Check standard certificate, instrument quality certificate, instruction manual, and HI7101412 rigid carrying case. Reagents sold separately | | |
| Reagents and Standards | HI97749 | HI97749-11 CAL Check standard cuvettes for chromium(VI) LR | |
| | | HI93749-01 chromium(VI) LR reagents for 100 tests | |
| | | HI93749-03 chromium(VI) LR reagents for 300 tests | |
| | | HI97723-11 CAL Check standard cuvettes for Chromium(VI) HR | |
| | HI97723 | HI93723-01 chromium(VI) HR reagents for 100 tests | |



HI93723-03 chromium(VI) HR reagents for 300 tests

<u>Photometers</u>

portable

10.58