

PM6902A

Formaldehyde Tester



PM6902B

Dust Particle Tester



www.tools.in.th
 ติดต่อบริษัท นีโอนิคส์ จำกัด
 Tel: 02-077-7602 หรือ 061-8268939
 E-mail: sale@tools.in.th หรือ sale@neonics.co.th

AMBIENT AIR QUALITY METER



Formaldehyde Tester

| Specifications | Measuring range | Resolving power | Accuracy |
|---------------------------------|--|--------------------------------|--------------------|
| HCHO | 0-10PPm, 0-13.4mg/m ³ | 0.01ppm, 0.01mg/m ³ | ±2% |
| Medium Sample Mode | Through built-in pump,collecting 10ml air sample | | |
| Sampling Frequency | 1-3 minute (depend on sample concentration) | | |
| Response time | Low concentration about 8 seconds, high concentration about 60 seconds | | |
| Display unit can be switched | ppm, mg/m ³ | | |
| Air Temperature Measuring range | -20-60°C | | ±0.5°C |
| Air humidity Measuring range | 0-100%RH | | ±2.5%RH(20%-80%RH) |
| Dew Point Temperature | -20°C-60°C | | |
| Wet Bulb Temperature | -20-60°C | | |

AMBIENT AIR QUALITY METER



Dust Particle Tester

| Function | Particle count, Temperature, Humidity, Dew point temperature |
|--|--|
| Particle weight/Particle channel | PM2.5, PM10 |
| Particle weight/Mass concentration range | 0-2000ng/m ³ |
| Particle weight/Display Resolution | 1ug/m ³ |
| Particle Count/Particle Channel | 0.3, 0.5, 1.0, 2.5, 5.0, 10um |
| Particle Count/Flow | Can be controlled through built-in sensor |
| Particle Count/Calculation Mode | Accumulation, difference, Concentration |
| Particle Count/Error | Accuracy : 5%, 2000000 particles/ft3; |
| Air Temperature /Measuring range | Range:-20-60°C ; Accuracy : ±0.5°C ; |
| Air humidity /measuring range | Range:0-100%RH ; Accuracy : ±2.5%RH(20%-80%RH) |
| Dew Point Temperature | Range:-20°C-60°C ; Accuracy : ±0.5°C ; |
| Wet Bulb Temperature | Range:-20-60°C ; Accuracy : ±0.5°C ; |
| Display Mode | Color LCD/LED Display |
| USD data export | Yes |
| Driving Power | 18650 2400mAH |